

1000 mL

RM/N 4057  
Rev. 100520

# Belzer MPS® (UW Machine Perfusion Solution)



**Warning: NOT intended for direct injection or I.V. infusion**  
**Warning: NOT for in situ flushing of organs in living donors or patients**  
**Warning: Discard any unused portions**

Osmolarity: 285–315 mosmol/kg  
pH: approximately 7.4

*Belzer MPS® is sterile, isotonic, non-pyrogenic and non-toxic. Belzer MPS® is formulated for the in-vitro flushing and continuous hypothermic machine perfusion preservation of explanted kidneys.*

*Belzer MPS® solution should be stored between 2° - 25°C (36° - 77°F). While 5°C is the ideal temperature for actual perfusion of Belzer MPS®, a range of 4° - 8°C is acceptable. Do not freeze or expose to excessive heat.*

**LOT:**

**EXP.:**

CAUTION: Federal (USA) law restricts this device to sale by or on the order of a physician.

**Manufactured by:**  
**Preservation Solutions, Inc.**  
1099 Proctor Drive  
Elkhorn, WI 53121

**Manufactured for:**  
**Bridge to Life Ltd.**  
128 Suber Road  
Columbia, SC 29210-3851

Belzer MPS® is a trademark of the Wisconsin Alumni Research Foundation (WARF) used under an exclusive license granted to Bridge to Life Ltd.